

WARNINGS

ActiPatch should not be used on persons with a cardiac pacemaker, cardioverter defibrillator, neurostimulator or any active medical or metallic implant that is within the area of application.

ActiPatch should not be used by women that are pregnant or think they may be pregnant.

ActiPatch should not be used by skeletally immature persons.

ActiPatch is not a sterile device. If using over an open wound, place the ActiPatch over sterile gauze.

Discontinue use if rash or skin irritation results. Consult a qualified health care practitioner before using ActiPatch again.

Do not stand, walk, or put body weight on ActiPatch as this may damage the circuitry.

ActiPatch can be disposed of in a normal trash container.

ActiPatch will not interfere with other electronic equipment in use.

Always consult a qualified health care practitioner to evaluate your condition or injury prior to using ActiPatch. If pain or swelling persist or worsen, discontinue use and seek medical attention immediately.

WARNINGS

ActiPatch should not be used on persons with a cardiac pacemaker, cardioverter defibrillator, neurostimulator or any active medical or metallic implant that is within the area of application.

ActiPatch should not be used by women that are pregnant or think they may be pregnant.

ActiPatch should not be used by skeletally immature persons.

ActiPatch is not a sterile device. If using over an open wound, place the ActiPatch over sterile gauze.

Discontinue use if rash or skin irritation results. Consult a qualified health care practitioner before using ActiPatch again.

Do not stand, walk, or put body weight on ActiPatch as this may damage the circuitry.

ActiPatch can be disposed of in a normal trash container.

ActiPatch will not interfere with other electronic equipment in use.

Always consult a qualified health care practitioner to evaluate your condition or injury prior to using ActiPatch. If pain or swelling persist or worsen, discontinue use and seek medical attention immediately.

U.S. & Foreign Patents Pending
Safe for Diabetics, Arthritics,
Elderly, the Bed Ridden and on Open Wounds



1-866-757-2284
Manufactured by:
BioElectronics Corporation
4539 Metropolitan Court
Frederick, MD 21704
Stock symbol: BIEL
www.BioElectronicsCorp.com



DFU-HEEL-USA-001

U.S. & Foreign Patents Pending
Safe for Diabetics, Arthritics,
Elderly, the Bed Ridden and on Open Wounds



1-866-757-2284
Manufactured by:
BioElectronics Corporation
4539 Metropolitan Court
Frederick, MD 21704
Stock symbol: BIEL
www.BioElectronicsCorp.com



DFU-HEEL-USA-001




Your ActiPatch Heel Pain Therapy Kit Contains:

- (1) 8cm Loop Device
- (1) Heel Wrap

The electronics and battery are in this module.

Therapeutic area is inside loop



The two most common causes of heel pain are Plantar Fasciitis and Achilles Tendonitis.


PLANTAR FASCIITIS is an inflammation of the tissue band connecting the heel bone to base of the toes. The pain is centered under the heel and usually flares up with the first steps in the morning after resting overnight.

ACHILLES TENDONITIS is an inflammation of the tendon that links your leg muscles to the back of your heel bone. This inflammation causes pain in the back of the heel and may cause the skin to thicken, get red and swell.

HOW TO USE ACTIPATCH TO TREAT PLANTAR FASCIITIS

Gently pull the tab to activate the ActiPatch. A red light indicates that the patch is on and working.

Position the loop of the ActiPatch on the bottom of the foot as illustrated below so that the inside of the loop is directly over the plantar fascia ligament. Secure the ActiPatch using the enclosed wrap. Do not over tighten the wrap as this may restrict circulation. You may also use medical grade tape.




For best results: Apply the ActiPatch immediately after injury or surgery and wear continuously until healed or for a minimum of 6-8 hours daily. Remove ActiPatch before bathing or showering and store in a cool dry location.

HOW TO USE ACTIPATCH TO TREAT ACHILLES TENDONITIS

Activate the ActiPatch.

Position the loop of the ActiPatch on back of the leg as illustrated below so that the inside of the loop is directly over the Achilles Tendon. Secure the ActiPatch using the enclosed wrap. Do not over tighten wrap as this may restrict circulation. You may also use medical grade tape.




Your ActiPatch Heel Pain Therapy Kit Contains:

- (1) 8cm Loop Device
- (1) Heel Wrap

The electronics and battery are in this module.

Therapeutic area is inside loop



The two most common causes of heel pain are Plantar Fasciitis and Achilles Tendonitis.


PLANTAR FASCIITIS is an inflammation of the tissue band connecting the heel bone to base of the toes. The pain is centered under the heel and usually flares up with the first steps in the morning after resting overnight.

ACHILLES TENDONITIS is an inflammation of the tendon that links your leg muscles to the back of your heel bone. This inflammation causes pain in the back of the heel and may cause the skin to thicken, get red and swell.

HOW TO USE ACTIPATCH TO TREAT PLANTAR FASCIITIS

Gently pull the tab to activate the ActiPatch. A red light indicates that the patch is on and working.

Position the loop of the ActiPatch on the bottom of the foot as illustrated below so that the inside of the loop is directly over the plantar fascia ligament. Secure the ActiPatch using the enclosed wrap. Do not over tighten the wrap as this may restrict circulation. You may also use medical grade tape.



For best results: Apply the ActiPatch immediately after injury or surgery and wear continuously until healed or for a minimum of 6-8 hours daily. Remove ActiPatch before bathing or showering and store in a cool dry location.

HOW TO USE ACTIPATCH TO TREAT ACHILLES TENDONITIS

Activate the ActiPatch.

Position the loop of the ActiPatch on back of the leg as illustrated below so that the inside of the loop is directly over the Achilles Tendon. Secure the ActiPatch using the enclosed wrap. Do not over tighten wrap as this may restrict circulation. You may also use medical grade tape.

